



**Bureau of Laboratory Quality Standards
Ministry of Public Health**

This is to certify that

The laboratory of

Foamtec International Co., Ltd

259/1, 259/2 Moo 3, Laemchabang Industrial Export

Processing Zone1, Toongsukhla,

Sriracha, Chonburi 20230, Thailand

has been accepted as an
accredited laboratory complying with the ISO/IEC 17025 : 2017
and the requirements of the Bureau of Laboratory Quality Standards

The laboratory has been accredited for specific tests
listed in the scope within the field of

Medical Devices Testing

A handwritten signature in blue ink that reads 'Patravee Soisangwan'.

(Dr. Patravee Soisangwan)

Director of Bureau of Laboratory Quality Standards

Date of Accreditation : 10 June 2022

Valid Until : 9 June 2026

Accreditation Number 1350/65

The laboratory of Foamtec International Co., Ltd has been accepted as an accredited laboratory in the field of medical devices testing for the following scopes.

No.	Type of Sample	Test	Method
1	Raw Material <ul style="list-style-type: none"> ● Polyurethane Foam ● Plastic Film 	1. <i>In Vitro</i> Cytotoxicity	ISO 10993-5: 2009
2	Medical Device Products <ul style="list-style-type: none"> ● Medical Foam Swab ● Nasopharyngeal Swab ● Oropharyngeal Swab ● Crimpers ● Mammography Pad Foam ● Polyester Wiper ● Polyurethane Foam Wiper 		
3	Medical Device Products <ul style="list-style-type: none"> ● Polyester Wiper 	2. Liquid Particle Count (LPC)	In-house method WCC-TM-LB-014 based on IDEMA Standard Microcontamination Document No.M9-98
		3. Non Volatile Residue (NVR)	In-house method WCC-TM-LB-019 based on IDEMA Standard Microcontamination Document No.M7-98